

|                        |                          |
|------------------------|--------------------------|
| Modello Laser          | <b>C-40A GEM-40A</b>     |
| Serial Number RF       | <b>ND</b>                |
| Serial Number Sorgente | <b>BTH.1138405.54719</b> |
| Sample Rate:           | <b>2000</b>              |
| Sample Period:         | <b>00:01,9</b>           |
| Total Duration:        | <b>1:02:19</b>           |
|                        | POWER MAX                |
|                        | SONDA + FILTRO H2O       |
|                        | CONDENSATORE 1000nF 63V  |

|                     |                         |
|---------------------|-------------------------|
| Data                | <b>08/05/2014 11:13</b> |
| Sonda               | <b>0164L12R</b>         |
| Frequenza           | <b>25 KHz</b>           |
| Duty                | <b>100 %</b>            |
| Valore Minimo       | <b>50 W</b>             |
| Valore Medio        | <b>52 W</b>             |
| Valore Massimo      | <b>53 W</b>             |
| Deviazione Standard | <b>0,9 W</b>            |
| Stabilità           | <b>3,62% +/-</b>        |

