

Modello Laser	E-400
Serial Number RF	4PA11370194
Serial Number Sorgente	ET53633
Sample Rate:	2000
Sample Period:	00.00.3
Total Duration:	0.11.23

Data	08/03/2013 15.54
Sonda	0164L12R
Frequenza	1 KHz
Duty	60 %
Valore Minimo	540 W
Valore Medio	558 W
Valore Massimo	577 W
Deviazione Standard	4.4 W
Stabilità	1.57% +/-

