

Modello Laser	E-400
Serial Number RF	EA3180
Serial Number Sorgente	ET54249
Sample Rate:	2000
Sample Period:	00:01,9
Total Duration:	1:02:20

Data	15/05/2013 10:37
Sonda	0164L12R
Frequenza	30 KHz
Duty	60 %
Valore Minimo	476 W
Valore Medio	506 W
Valore Massimo	525 W
Deviazione Standard	13,5 W
Stabilità	5,35% +/-

