

Modello Laser	<b>E-400</b>
Serial Number RF	<b>EA3178</b>
Serial Number Sorgente	<b>ET54327</b>
Sample Rate:	<b>2000</b>
Sample Period:	<b>00:00,3</b>
Total Duration:	<b>0:11:20</b>

Data	<b>11/06/2013 11:08</b>
Sonda	<b>0164L12R</b>
Frequenza	<b>1 KHz</b>
Duty	<b>60 %</b>
Valore Minimo	<b>550 W</b>
Valore Medio	<b>579 W</b>
Valore Massimo	<b>597 W</b>
Deviazione Standard	<b>14,4 W</b>
Stabilità	<b>4,99% +/-</b>

