

Modello Laser	K-250
Serial Number RF	135MQ1202008
Serial Number Sorgente	HB47154
Sample Rate:	2000
Sample Period:	00:00,3
Total Duration:	0:11:19

Data	18/07/2013 11:41
Sonda	0164L12R
Frequenza	1 KHz
Duty	60 %
Valore Minimo	330 W
Valore Medio	338 W
Valore Massimo	348 W
Deviazione Standard	4,0 W
Stabilità	2,37% +/-

